



(12) **United States Patent**  
**Parsai et al.**

(10) **Patent No.:** **US 9,370,672 B2**  
(45) **Date of Patent:** **Jun. 21, 2016**

(54) **UNFILTERED RADIATION THERAPY**

(75) Inventors: **E. Ishmael Parsai**, Ottawa Hills, OH (US); **John J. Feldmeier**, Monroe, MI (US)

(73) Assignee: **The University of Toledo**, Toledo, OH (US)

(\* ) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 1663 days.

(21) Appl. No.: **12/224,015**

(22) PCT Filed: **Feb. 20, 2007**

(86) PCT No.: **PCT/US2007/004403**

§ 371 (c)(1), (2), (4) Date: **Oct. 13, 2010**

(87) PCT Pub. No.: **WO2007/098164**

PCT Pub. Date: **Aug. 30, 2007**

(65) **Prior Publication Data**

US 2011/0033028 A1 Feb. 10, 2011

**Related U.S. Application Data**

(60) Provisional application No. 60/775,677, filed on Feb. 21, 2006.

(51) **Int. Cl.**  
**A61N 5/10** (2006.01)

(52) **U.S. Cl.**  
CPC ..... **A61N 5/1042** (2013.01); **A61N 5/1077** (2013.01); **A61N 2005/1034** (2013.01)

(58) **Field of Classification Search**  
CPC ... **A61N 5/103**; **A61N 5/1031**; **A61N 5/1036**;  
**A61N 5/1042**; **A61N 5/1045**; **A61N 5/1047**;  
**A61N 5/1077**

USPC ..... **378/65**  
See application file for complete search history.

(56) **References Cited**

**U.S. PATENT DOCUMENTS**

4,868,843 A \* 9/1989 Nunan ..... 378/152  
5,216,255 A \* 6/1993 Weidlich ..... 250/492.3

(Continued)

**FOREIGN PATENT DOCUMENTS**

DE WO9429882 12/1994

**OTHER PUBLICATIONS**

Weihua Fu et al.; "Delivery time comparison for intensity-modulated radiation therapy with/without flattening filter: a planning study; IMRT delivery time comparison with/without flattening filter"; Physics in Medicine and Biology, Institute of Physics Publishing, Bristol GB, vol. 49, No. 8, Apr. 21, 2004, pp. 1535-1547, XP020024089, ISSN: 0031-9155, DOI: 10.1088/0031-9155/49/8/011.

(Continued)



*Primary Examiner* — Thomas R Artman

(74) *Attorney, Agent, or Firm* — Emch, Schaffer, Schaub & Porcello Co., L.P.A.

(57) **ABSTRACT**

This is a new technique in IMRT and 3D conformal gamma radiation dose delivery using a linear accelerator with no flattening filter. The technique improves patient radiation therapy by reducing radiation scattered to surrounding normal tissue and reducing electron contamination. It increases dose rate to shorten treatment time. Linear accelerators have for decades come with a photon flattening filter to make the photon profile of planar fluence to make the dose distribution more uniform. These filters, however, resulted in fluence attenuation and contamination of the beam. Now in the age of techniques such as intensity modulated radiation therapy (IMRT) the function of the flattening filter becomes redundant. The flattening filter now merely reduces the efficiency of the beam by reducing the fluence and increasing scattered radiation. Our technique involves removal of the flattening filter for complex treatments. It uses inverse planning along with multi-leaf collimators to shape the dose distribution.

**2 Claims, 8 Drawing Sheets**



(56)

**References Cited**

U.S. PATENT DOCUMENTS

5,332,908 A \* 7/1994 Weidlich ..... 250/492.1  
 5,596,619 A \* 1/1997 Carol ..... 378/65  
 5,668,371 A \* 9/1997 Deasy et al. .... 850/1  
 5,802,136 A \* 9/1998 Carol ..... 378/65  
 6,038,283 A \* 3/2000 Carol et al. .... 378/65  
 6,393,096 B1 \* 5/2002 Carol et al. .... 378/65  
 7,162,008 B2 1/2007 Earl et al.  
 7,295,649 B2 11/2007 Johnsen  
 7,400,434 B2 \* 7/2008 Brahme et al. .... 359/237  
 7,945,022 B2 \* 5/2011 Nelms et al. .... 378/65  
 7,961,843 B2 \* 6/2011 Brown et al. .... 378/65  
 8,027,431 B2 \* 9/2011 Stahl et al. .... 378/65

8,050,384 B2 \* 11/2011 Carol et al. .... 378/65  
 2003/0086527 A1 5/2003 Speiser et al.  
 2004/0071261 A1 \* 4/2004 Earl et al. .... 378/65  
 2004/0082855 A1 \* 4/2004 Robar et al. .... 600/436  
 2007/0086569 A1 \* 4/2007 Johnsen ..... 378/65  
 2008/0049897 A1 \* 2/2008 Molloy ..... 378/65  
 2008/0170663 A1 \* 7/2008 Urano et al. .... 378/65  
 2008/0191141 A1 \* 8/2008 Nilsson ..... 250/393  
 2010/0305905 A1 \* 12/2010 Ehringfeld ..... 702/189  
 2011/0033028 A1 \* 2/2011 Parsai et al. .... 378/65

OTHER PUBLICATIONS

Extended European Search Report; Dated Sep. 25, 2013.

\* cited by examiner




Fig. 1a




Fig. 1b




Fig. 2a




Fig. 2b




Fig. 3




Fig. 4




Fig. 5a




Fig. 5b




Fig. 6a




Fig. 6b






Fig. 7a




Fig. 7b




Fig. 8

## UNFILTERED RADIATION THERAPY

## CROSS REFERENCE TO RELATED APPLICATIONS

This application claims the benefit of U.S. provisional patent application: Ser. No. 60/775,677 filed Feb. 21, 2006.

## TECHNICAL FIELD

This invention relates to a method of performing radiation therapy. More specifically the invention relates to a new technique in IMRT conformal gamma radiation dose delivery using a linear accelerator with no flattening filter. The new technique improves patient radiation therapy by reducing radiation scattered to surrounding normal tissue without a filter.

## BACKGROUND OF THE INVENTION

Intensity modulated radiation therapy (IMRT) is a treatment method for cancer patients requiring radiation treatment. IMRT is an extremely precise method of treatment delivery where the radiation dose conforms to the target and avoids the surrounding critical structures. Rather than having a single large radiation beam pass through the body, with IMRT the treatment is delivered from various angles and the intensity of the radiation beam is varied across the treatment area.

The radiation is effectively broken up into thousands of tiny pencil-thin radiation beams. With millimeter accuracy, these beams enter the body from many angles and intersect on the cancer. This results in a high radiation dosage to the tumor and a lower radiation dose to the surrounding healthy tissues.

One method for modulating the intensity of the radiation beam is based upon moving a multi-leaf collimator (MLC) in and out of radiation beam from the radiation treatment machine. An MLC comprises a plurality of thin width mechanical blades or leaves, which are individually controlled by miniature motors and mechanical drive linkages. A computer controls the miniature motors for driving the individual blades in and out to shape the radiation beam. An advantage of an MLC based IMRT treatment machine is that the same MLC can be automatically controlled to support the individual needs of each patient receiving radiation treatment. In other words, the MLC is reconfigured for each new patient.

Linear accelerators have for decades come with a photon flattening filter to make the photon profile of planar fluence and thus, the dose distribution more uniform. These filters have then resulted in fluence attenuation and contamination of the beam. Now in the age of techniques such as intensity modulated radiation therapy (IMRT) the function of the flattening filter becomes redundant and the flattening filter now merely reduce the efficiency of the beam by reducing the fluence and increase scattered radiation.

Other objects and advantages of the present invention will become apparent to those skilled in the art upon a review of the following detailed description of the preferred embodiments and the accompanying drawings.

## SUMMARY OF THE INVENTION

Our technique involves removal of the flattening filter for complex treatments and using inverse planning along with multi-leaf collimators to shape the dose distribution.

With the flattening filter removed the dose rate is increased and the lateral scatter is reduced. This improves patient treatment by reducing dose to the normal tissue surrounding the target and also reduces treatment times. The flattening of the beam profile is redundant in techniques such as IMRT since the planar fluence is controlled by the multi-leaf collimator (MLC). For many modern linear accelerators, removal of the flattening filter requires no physical modification of the unit since the flattening filter can simply be mechanically moved out of the beam path.

This new technique is in IMRT and 3D conformal gamma radiation does delivery using a linear accelerator with no flattening filter. The technique improves patient radiation therapy by reducing radiation scattered to surrounding normal tissue and reducing electron contamination. It increases dose rate to shorten treatment time.

Linear accelerators have for decades come with a photon flattening filter to make the photon profile of planar fluence to make the dose distribution more uniform. These filters, however, have resulted in fluence attenuation and contamination of the beam.

Now in the age of techniques such as intensity modulated radiation therapy (IMRT) the function of the flattening filter becomes redundant. The flattening filter now merely reduces the efficiency of the beam by reducing the fluence and increasing scattered radiation.

## BRIEF DESCRIPTION OF THE DRAWINGS

FIGS. 1(a) and 1(b) show Monte Carlo and water phantom measurements of the CAX percent depth-dose for 6 MV and 10 MV.

FIGS. 2(a) and 2(b) show Monte Carlo and water phantom measurements of transverse profiles at depths  $d_{max}$  and 10 cm, for 6 MV and 10 MV.

FIG. 3 shows a comparison between Monte Carlo and measured data for a 6 MV  $10 \times 10$  cm<sup>2</sup> beam.

FIG. 4 shows a Monte Carlo computed transverse cross-plane profiles at a depth of 1.6 cm for a 6 MV filter free photon beam of field size ranging from  $2 \times 2$  to  $30 \times 30$  cm<sup>2</sup>.

FIGS. 5(a) and 5(b) shows a comparison between Monte Carlo simulations for a standard, flattened and an unflattened 6 MV and 10 MV  $10 \times 10$  cm<sub>2</sub> beam at  $d_{max}$ . All profiles are normalized to the central axis dose of the standard beam to show the effect on the CAX dose of removing the flattening filter.

FIGS. 6(a) and 6(b) shows a Monte Carlo percent depth-dose curves comparing the standard flattened 6 MV and 10 MV beams to the equivalent filter-free beams.

FIGS. 7(a) and 7(b) show a photon fluence spectra for a 6 MV beam and a 10 MV beam showing the effect of removing the photon flattening filter.

FIG. 8 shows a comparison of absolute doses obtained from simulations of  $2 \times 2$ ,  $10 \times 10$  and  $30 \times 30$  cm<sup>2</sup> fields. The simulations shown here were for a 6 MV beam at a depth of 1.6 cm.

## DETAILED DESCRIPTION OF THE INVENTION

IMRT (Intensity Modulated Radiation Therapy) is rapidly becoming a common treatment modality with a recent study claiming that it is used by a third of the radiation oncologists in the United States. The modern treatment machines are designed with dynamic MLC and IMRT-ready systems integrated into them but many of the current linear accelerators still used today, have the MLC as an add-on. In either case the linear accelerator is designed such that IMRT treatments and standard treatments can be carried out on the same unit. The

conventional 3D conformal therapy treatment requires a flat beam because generally dose compensation to achieve uniformity within target volume for each individual beam is not performed. However, in generating IMRT treatment plans, the planner ends up with a non-uniform density matrix to deliver the desired dose with the target volume, and spare the surrounding normal or critical structures. To achieve this goal, a flat beam is not required. Modulation of beam during IMRT planning and delivery is performed through segmented fields and many beamlets within the delivery port and in fact thinking out side of convention, one would see the advantages in having a cleaner beam that does not need to suffer all the scattering through a thick chunk of metal, namely the flattening filter. It is therefore, expected that removal of the flattening filter would lead to better IMRT treatments due to the reduction in lateral photon scatter and the increase in central axis photon fluences. More specifically, by moving the flattening filter out of the path of the beam solely for IMRT treatments, higher dose rates and sharper, more geometrically defined fields can be expected thus leading to better IMRT plans and treatments. The deleterious properties of the flattening filter are caused by the increased lateral scatter and the decreased central axis fluence that the filter produces. In the special case of IMRT, where fluence is varied by a combination of MLC movements and beam modulation at the patient level, the filter is no longer required. This has been shown for the specific case of tomotherapy; a dedicated IMRT system. Here we show Monte Carlo simulations of radiation characteristics for the more general case of a linear accelerator.

#### Methods and Materials

##### Monte Carlo Simulations of an Elektra SL-25

Monte Carlo simulations were carried out using the BEAMnrc code. Using an Elekta precise model SL-25 photon beams of 6 MV and 10 MV energies were initially modeled and commissioned by comparing the simulations to data measured using a Wellhofer (Scanditronix Wellhofer) scanning water phantom. In the models the head of the accelerator was broken down into component modules, namely the target, primary collimator, flattening filter, monitor chamber, mirror, MLC and X and Y jaws. An additional component was used to simulate the air gap between the exit of the accelerator and the water phantom surface, where the phase space plane was located. The energy cutoffs for transport were set as ECUT=0.7 MeV, PCUT=0.01 MeV and global electron cutoff=2.0 MeV. Electron range rejection and selective Bremsstrahlung splitting were used, with SBS parameters  $N_{min}=10$  and  $N_{max}=100$ . Russian roulette and photon forcing were not employed. The phase space file created at the plane 100 cm from the source was then used as the input for the phantom, simulated using the DOSXYZarc code.  $400 \times 10^6$  histories were used for the simulation of the accelerator. For the DOSXYZ phantom,  $200 \times 10^6$  histories were used for all field sizes, resulting in adequate statistics for the larger field sizes.

Both depth dose and transverse profiles depend greatly on the properties of the electron beam as it strikes the photon target. The parameters of importance are the mean electron energy, the energy spread and the spatial distribution of the beam. For the 6 MV and 10 MV beams respectively, the electron energy used was 6.50 MeV and 9.50 MeV, the energy spread was 1.0 MeV and 0.8 MeV FWHM and the radial distribution was 0.11 cm and 0.10 cm FWHM. Depth dose curves obtained from these simulations deviated less than 1% in the region of dose-maximum and less than 5% at all other depths, when compared to water phantom measurements.

Once the Monte Carlo simulation was found to match the measured data to adequate levels, the flattening filters were

removed from both of the 6 MV and 10 MV beam models. All other parameters remained unchanged. Measurements Made in Water.

All measurements were made at 100 cm SSD in a Wellhofer scanning water phantom, with a 0.1 cc ionization chamber. Both 6 MV and 10 MV beams were studied for comparison with the Monte Carlo simulations. After the Monte Carlo model commissioning data was obtained the 6 MV and 10 MV flattening filters were removed from the primary rotating carousel in the head of the accelerator. This left a hole in the carousel which the photon beam could pass through. Depth-ionization profiles, transverse inline (gun-target direction) and cross-plane profiles were measured at  $d_{max}$  and 10 cm. Depth dose profiles were measured to a depth of 30 cm and normalized to the maximum chamber reading on the central axis. Transverse profiles were measured in the inline and cross-plane directions for field sizes ranging from  $5 \times 5$  to  $30 \times 30$  cm<sup>2</sup>. These profiles were also normalized to the maximum chamber reading on the central axis.

#### Results

##### Monte Carlo Model Commissioning

As mentioned, Monte Carlo simulations of the standard, flattened 6 MV and 10 MV beams were carried and they matched well with the measured data obtained with the scanning water phantom. The purpose of these measurements was to show that the Monte Carlo models accurately match the measurements of dose performed in the water phantom.

FIGS. 1(a) and 1(b) show Monte Carlo and water phantom measurements of the CAX percent depth-dose for 6 MV and 10 MV. Central axis percent depth-dose profiles for a  $10 \times 10$  cm<sup>2</sup> field at 100 cm SSD are shown for 6 MV and 10 MV, with the experimental measurements shown as solid points and the Monte Carlo model shown as hollow points. Transverse profiles of a  $30 \times 30$  cm<sup>2</sup> field were also obtained for comparison of the flatness and symmetry of the Monte Carlo models with respect to the measured data.

FIGS. 2(a) and 2(b) show 6 MV and 10 MV Monte Carlo calculated transverse profiles for the inline direction at depths of  $d_{max}$  and 10 cm, compared to the measured data. A good agreement between measured and Monte Carlo modeled data was found in all cases.

FIG. 3 shows a comparison between Monte Carlo and measured data for a 6 MV  $10 \times 10$  cm<sup>2</sup> beam. The top two curves are for a depth of 1.6 cm ( $d_{max}$ ) and the bottom two curves are for a depth of 10 cm.

##### Monte Carlo Modelling of a Non-Flat Beam

Simulations were then carried out without the filter and compared to data measured after the flattening filters had been removed from the primary filter carousel of the Elekta accelerator. The purpose of these measurements was to verify the accuracy of the Monte Carlo models to accurately simulate a beam without the flattening filter. A comparison for the cross-plane profiles is shown in FIG. 3.

Not shown are the comparisons between the inline (gun-target) direction measured and Monte Carlo profiles. These measured transverse profiles had poor symmetry and this was believed to be due to difficulties of steering the beam after removal of the flattening filter. It can be concluded from FIG. 3 that the Monte Carlo models of filter free 6 MV and 10 MV beams were shown to accurately match the measured data. Simulations were then carried out for various field sizes ranging from  $2 \times 2$  cm<sup>2</sup> to  $30 \times 30$  cm<sup>2</sup>. The graph below shows the transverse profiles obtained at 1.6 cm depth for a 6 MV beam without a flattening filter. The curves in FIG. 4 are all normalized to the CAX dose of the  $10 \times 10$  cm<sup>2</sup> field.

## 5

FIG. 4 shows Monte Carlo computed transverse cross-plane profiles at a depth of 1.6 cm for a 6 MV filter free photon beam of field size ranging from 2×2 to 30×30 cm<sup>2</sup>.

The next step was to compare the Monte Carlo models of the flattened and unflattened beams. FIG. 5 shows Monte Carlo calculated transverse profiles and the effect on the central axis (CAX) dose of removing the flattening filter. It was found that for the 6 MV photon beam of 10×10 cm<sup>2</sup> field size the CAX dose was increased by a factor of 2.35 with the filter removed, compared to the standard flattened beam. This figure also shows the CAX dose for a 10×10 cm<sup>2</sup> 10 MV beam with and without the flattening filter. In this case, since the 10 MV flattening filter for the Elekta is more substantial in terms of mass of material used the CAX dose without the filter is 4.18 times higher than the standard flattened beam.

FIGS. 5(a) and 5(b) show a comparison between Monte Carlo simulations for a standard, flattened and a unflattened 6 MV and 10 MV 10×10 cm<sup>2</sup> beam at  $d_{max}$ . All profiles are normalized to the central axis dose of the standard beam to show the effect on the CAX dose of removing the flattening filter.

### C. Quantification of Beam Flatness

The flatness of each transverse profile was calculated using the variation over the mean at 80% of the field size, with the equation,

$$\text{flatness} = 100 \times \frac{D_{max} - D_{min}}{D_{max} + D_{min}}$$

For the 6 MV simulation of a 10×10 cm<sup>2</sup> beam, the flatness at  $d_{max}$  was 2.37% and 6.21% for the flattened and unflattened beam, respectively. Similarly, at 10 cm depth the equivalent percentages were 1.88% and 5.77%.

For the 10 MV simulations, flatness percentages of 3.96% and 7.71% were obtained at depths of 2.3 cm ( $d_{max}$ ) and 10 cm. for the standard and unflattened beam, respectively. At 10 cm depth flatness was calculated to be 2.92% for the flattened beam and 8.39% for the unflattened beam.

### D. Dose on the Central Axis

FIGS. 6(a) and 6(b) show Monte Carlo percent depth-dose curves comparing the standard flattened 6 MV and 10 MV beams to the equivalent filter-free beams. The faster decrease in dose with depth for the filter-free beam is consistent with a softer central axis beam.

Depth dose curves on the central axis were also obtained from simulations of the flattened and unflattened 10×10 cm<sup>2</sup> 6 MV and 10 MV beams. The dose deposited at depths greater than  $d_{max}$  was found to decrease more rapidly with the filter removed. This is due to the fact that, with the filter removed the beam in the region of the central axis is no longer hardened by the filter. The faster decrease in dose with depth is consistent with a softer central axis beam. To investigate the effect of the flattening filter on the photon energy spectrum an analysis of various phase space files with the program BEAMDP was performed. Photon fluence as a function of photon energy was graphed for the filter free beams versus the standard beams. As expected, the photon fluence per unit energy is significantly greater for the filter free beam, especially in the region of the peak photon energy.

FIGS. 7(a) and 7(b) show photon fluences spectra for a 6 MV beam and a 10 MV beam showing the effect of removing the photon flattening filter. FIG. 7 shows the photon fluence spectra across a 10×10 cm<sup>2</sup> field for both the 6 MV and the 10 MV beam. In both cases the peak photon energy is increased by removing the flattening filter, showing that the flattening

## 6

filter has the effect of hardening the beam. For the 6 MV beam the peak energy with and without the flattening filter are 0.48 MV and 0.33 MeV respectively. Similarly, for the case of the 10 MV beam, where the design of the flattening filter leads to a greater beam hardening effect, the peak photon energies are 1.13 MeV and 0.33 MeV for the standard beam and the filter free beam.

### E. Out of Field Dose

With the flattening filter removed, one would expect the amount of lateral photon scatter to decrease, the effect being that the dose at a point outside the field would be reduced. To investigate this effect a comparison between the relative dose at and beyond the edge of the radiation field was made between simulations made of a 6 MV beam with and without the flattening filter. Simulations were run for a 6 MV beam for various field sizes ranging from 2×2 cm<sup>2</sup> to 30×30 cm<sup>2</sup>. In all cases the dose at the edge of the field was greater for the filter-free beam. In FIG. 8 below, 2×2 cm<sup>2</sup>, 10×10 cm<sup>2</sup> and 30×30 cm<sup>2</sup> fields are shown for a flattened and filter-free 6 MV beam. It can be seen that, in the wings of the profile the relative dose for the filter-free beam is greater than that of the standard field in all cases. The profiles below are at a depth of 1.6 cm. The same profiles at a depth of 10 cm showed the same effect; the out of field dose being higher for the filter free beam.

FIG. 8 is a comparison of absolute dose obtained from simulations of 2×2, 10×10 and 30×30 cm<sup>2</sup> fields. The simulations shown here were for a 6 MV beam at a depth of 1.6 cm. For each field size a profile of the flattened beam and the unflattened beam are shown so that the dose at the edges of the radiation field can be compared. It can be seen that for all field sizes the dose at the edge of the field is greater for the filter-free beam.

To quantify the out of field dose we considered a point 2 cm outside of the field (e.g. at an off axis distance of 3 cm for a 2×2 cm<sup>2</sup> field) and took the average of the relative doses for the voxels to the right and left of the central axis. The table below shows the relative dose (the normalization is with respect to the CAX dose for the standard, flattened beam for that field size) at a point 2 cm outside the radiation field, for both the flattened and unflattened 6 MV beams. All profiles considered here are at a depth of maximum dose.

TABLE 1

	Relative Dose (0%)	
	with filter	No filter
2 × 2 cm <sup>2</sup>	0.63	1.10
10 × 10 cm <sup>2</sup>	2.86	4.55
30 × 30 cm <sup>2</sup>	6.03	7.64

### Conclusions:

With flattening filter removed, the photon beams will not suffer the remarkable scattering that they will go through otherwise, resulting in a much cleaner beam at the patient's level. The conventional treatments requiring a flat photon beam are not necessary for IMRT treatments as the beams are modulated to achieve dose uniformity within the target volume. In fact the fluence maps as generated from a beam end up being very nonuniform for IMRT cases. The substantial increase in dose rate from a flattening filter free accelerator is significant in delivering a less contaminated beam at much shorter times. The computed depth-dose plots for both 6 and

10 MV photon beams indicate that by removing the flattening filter out of the beam, better dose fall off beyond depth of maximum dose is achieved. On the other hand, because of a less hardened beam, the point of maximum dose at depth will get closer to the surface (1-2 mm for 6x, and 2-3 mm for 10x). The out of field dose is a phenomenon that requires further study and will be discussed in detail in future works, but the measured and computed dose profiles in treatment fields indicate less scatter, significantly higher photon fluence, and overall a cleaner beam to be used for the IMRT treatment. The better fall-off of the dose beyond depth of maximum dose in a flattening free accelerator is also another indication to cleaner beams when filter is removed. The quantities of scatter and lower energy photons contributing to dose at depth is directly proportional to the energy of the beam and is considerable for clinical photon beams.

#### Modifications

Specific compositions, methods, or embodiments discussed are intended to be only illustrative of the invention disclosed by this specification. Variation on these compositions, methods, or embodiments are readily apparent to a person of skill in the art based upon the teachings of this specification and are therefore intended to be included as part of the inventions disclosed herein.

The above detailed description of the present invention is given for explanatory purposes. It will be apparent to those skilled in the art that numerous changes and modifications

can be made without departing from the scope of the invention. Accordingly, the whole of the foregoing description is to be construed in an illustrative and not a limitative sense, the scope of the invention being defined solely by the appended claims.

We claim:

1. A radiation therapy device comprising:
  - a radiation source that directs an unmodulated raw beam along a beam path toward a selected treatment zone; wherein the radiation source is a linear accelerator without a flattening filter to provide a high radiation dose; wherein the radiation dose is high energy X-rays ranging from 6 MV to 25 MV;
  - a beam shaping device controllable to selectively collimate the unmodulated raw beam including at least one multi-leaf collimators (MLC); and
  - a treatment planning system for delivering an Intensity Modulated Radiation Therapy (IMRT) by utilizing the raw unmodulated beam having a predetermined profile of the high energy radiation dose and using the MLC to shape and modulate the raw beam profile for a particular dose distribution, wherein by not using a flattening filter, the MLC is configured to apply a predetermined modulation of the high energy radiation dose to be applied to the predetermined treatment zone.
2. The radiation device of claim 1, wherein the modulated beam has a higher photon fluence compared to a flattened beam.

\* \* \* \* \*